

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100406

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 03/10/2012

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 27.90: 0.074
LOT #: 02511080A1
EXPIRATION: 02/01/2013

BLANK TEST	0.000	15:20
INTERNAL STANDARD	VERIFIED	15:21
EXTERNAL STANDARD	0.073	15:21
BLANK TEST	0.000	15:22
EXTERNAL STANDARD	0.073	15:22
BLANK TEST	0.000	15:23
EXTERNAL STANDARD	0.073	15:23
BLANK TEST	0.000	15:24
EXTERNAL STANDARD	0.073	15:24
BLANK TEST	0.000	15:25
EXTERNAL STANDARD	0.073	15:25
BLANK TEST	0.000	15:26

Average = 0.0730
Std Dev = 0.0000

Diagnostic Check

VERSIONS

DMT: 2.00
PIC: 3.02
Modem: 2.3
Questions: 2.2

TEMPERATURES

Sample Chamber = 48.7°C	PASSED
Breath Tube = 46.3°C	PASSED

PUMP INFO

Flow Rate = 4.679 L/M	PASSED
-----------------------	--------

DETECTOR INFO

PUMP ON	PASSED
PUMP OFF	PASSED


FILTER INFO

Filter 1	PASSED
Filter 2	PASSED
Filter 3	PASSED

INTERNAL STANDARD	PASSED
-------------------	--------


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the Breath Alcohol Program, which includes assuring that instruments are calibrated and maintaining program records.
- (5) The above is a true and accurate verification of calibration, which is performed by the instrument's software, as specified by the State Breath Alcohol Program. Verification of calibration is a regularly conducted and regularly recorded activity of the State Breath Alcohol Program.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.


Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 6 day of April, 20 12


Nikki Roth, Notary Public
My Commission Expires With Office




2/14/12